



REPAIR REQUEST

Medical Maintenance Operations Division

TRACY

PART I – CUSTOMER DATA			PART II – MAINTENANCE ACTIVITY DATA		
1a. UIC	1b. Unit Name	1c. Phone	4a. Activity	4b. UIC	4c. DODAAC
		1d. DSN:	MMOD-CA	W05J02	W62SEV
2a. DODAAC	2b. POC Phone	2c. POC Email	5a. POC Administrator	5b. POC Phone 209-839-4557	5c. DSN 462-4557
3a. Alternate	3b. Alt. Phone	3c. Alt. Email	6a. Alternate Product Control	6b. Alt. Phone 209-839-5578	6c. Alt. DSN 462-5578
3d. Owing Organization UIC:			6c. Email usarmy.detrick.usamma.mesg.tracy-supervisors@army.mil		
EQUIPMENT RETURN TO ADDRESS – CUSTOMER			SHIP TO ADDRESS – MAINTENANCE ACTIVITY		
			ATTN: Production Control US ARMY MEDICAL MATERIEL AGENCY MEDICAL MAINTENANCE – CA 25600 SOUTH CHRISMAN ROAD WHSE 6 TRACY, CA 95304-9150		
DODAAC:		DODAAC <i>Same as above</i>			
PART III – CUSTOMER OPTIONS					
Email update		Tele Maintenance Request	Request EXCHANGE (IF AVAILABLE) Property Book transaction. Serial Number change.		
PART IV – EQUIPMENT LIST					
NOTE: Per CDC, contaminated equipment must be disinfected with a 1:10 – 1:100 dilution of 5.25 – 6.15% sodium hypochlorite (bleach: water).					
GCCS Work Order	NSN	Nomenclature	Serial Number	Symptom or Fault	
ADDITIONAL INSTRUCTIONS: **PLEASE REMEMBER TO INCLUDE THE GCCS-ARMY WORK ORDER FOR ALL DEVICES** Please provide the following for each equipment item (if possible): 1. All accessories (power cord, ECG cables, etc.) 2. Manufacturer service literature for non-standard equipment. If the equipment cannot be repaired, it will be condition coded IAW AR 725-50. The owning or supporting unit will be notified for disposition of device(s). SHIPPING INSTRUCTIONS: 1. Properly pack the equipment to ensure it is not damaged in transport. 2. Include this form with the device(s) or email it to the maintenance activity.					



REPAIR REQUEST

Medical Maintenance Operations Division

TRACY

GCCS Work Order	NSN	Nomenclature	Serial Number	Symptom or Fault

ADDITIONAL INSTRUCTIONS: **PLEASE REMEMBER TO INCLUDE THE GCSS-ARMY WORK ORDER FOR ALL DEVICES**

Please provide the following for each equipment item (if possible):

- 1. All accessories (power cord, ECG cables, etc.)
- 2. Manufacturer service literature for non-standard equipment.

If the equipment cannot be repaired, it will be condition coded IAW AR 725-50. The owning or supporting unit will be notified for disposition of device(s).

SHIPPING INSTRUCTIONS:

- 1. Properly pack the equipment to ensure it is not damaged in transport.
- 2. Include this form with the device(s) or email it to the maintenance activity.

This form may be reproduced but shall contain the same information as it appears on this form.