



REPAIR & RETURN

Maintenance Request

Medical Maintenance Operations Division-Hill AFB, UT

R&R HILL

PART I – CUSTOMER DATA			PART II – MAINTENANCE ACTIVITY DATA		
1a. UIC	1b. UNIT NAME	1c. PHONE NO	4a. UIC W05J08	4b. DODAAC W81PYK	4c. ACTIVITY NAME Medical Maintenance Operations Division - UT
2a. DODAAC	2b. POC	2c. DSN	5a. POC Administrator	5b. PHONE 801-586-5093	5c. DSN 586-5093
3a. EMAIL	3b. Owning Organization UIC		6a. 2 nd POC Production	6b. PHONE 801-586-4949	6c. EMAIL MCMR.USAMMA-MMOD-UT.ProCon@us.af.mil
EQUIPMENT RETURN TO ADDRESS - CUSTOMER			EQUIPMENT SHIP TO ADDRESS – MAINTENANCE ACTIVITY		
DODAAC: DODAAC Same as above			ATTN: Production Control US ARMY MEDICAL MATERIEL AGENCY MEDICAL MAINTENANCE – UT 6149 WARDLEIGH ROAD BLDG. 1160, BAY 1 HILL AFB, UT 84056-5848		
PART III – CUSTOMER OPTIONS					
Request email updates		Tele Maintenance Request		Request EXCHANGE (IF AVAILABLE) Property Book transaction. Serial number change.	
PART IV – EQUIPMENT LIST					
NOTE: Per CDC, contaminated equipment must be disinfected with a 1:10–1:100 dilution of 5.25%–6.15% sodium hypochlorite (bleach: water).					
GCSS Work Order #	NSN	NOMENCLATURE	SERIAL NUMBER	SYMPTOM/FAULT	
ADDITIONAL INSTRUCTIONS: **PLEASE REMEMBER TO INCLUDE THE GCSS-ARMY WORK ORDER FOR ALL DEVICES**					
Please provide the following for each equipment item (whenever possible):					
1. All accessories (power cord, ECG cables, etc.).					
2. Manufacturer service literature for non-standard equipment (if available).					
If the equipment is not repairable it will be condition coded IAW AR 725-50. The owning or supporting unit will be notified to determine the disposition of the device(s).					
SHIPPING INSTRUCTIONS:					
1. Properly pack the equipment to ensure it is not damaged in shipping.					
2. Attach this form to each shipping container.					

For questions, please call: 801-586-5093, DSN: 586-5093



REPAIR & RETURN

Maintenance Request

Medical Maintenance Operations Division-Hill AFB, UT

R&R

HILL

GCSS Work Order #	NSN	NOMENCLATURE	SERIAL NUMBER	SYMPTOM/FAULT

ADDITIONAL INSTRUCTIONS: **PLEASE REMEMBER TO INCLUDE THE GCSS-ARMY WORK ORDER FOR ALL DEVICES**
Please provide the following for each equipment item (whenever possible):
1. All accessories (power cord, ECG cables, etc.).
2. Manufacturer service literature for non-standard equipment (if available).
If the equipment is not repairable it will be condition coded IAW AR 725-50. The owning or supporting unit will be notified to determine the disposition of the device(s).
SHIPPING INSTRUCTIONS:
1. Properly pack the equipment to ensure it is not damaged in shipping.
2. Attach this form to each shipping container.

For questions, please call: 801-586-5093, DSN: 586-5093

This form may be reproduced but shall contain the same information as it appears on this form.